

Instruction Manual

HAND-HELD ELECTROSTATIC VOLTMETER ARS-H002ZA



* Read this instruction manual before using the product in order to achieve maximum performance.

% Keep this instruction manual within your reach after reading so that it can be used at any time.% Warranty is on the last page of this manual. Please read the contents and keep it carefully.

28, Namyang-ro 930 beon-gil, Hwaseong-si, Gyeonggi-do, Korea Tel 82.31.299.5453 Fax 82.31.357.2610

CONTENTS

1. Features		3 Page
2. Specification		4 Page
3. Package contents		5 Page
4. Parts name and functions .		6 Page
5. Dimension		8 Page
6. Performance		9 Page
7. Operating Instruction		10 Page
7.1 Power supply[Battery use]	7.5 Ion Balance mode	
7.2 Measuring set-up	7.6 HOLD function	
7.3 Setting zero point	7.7 Automatic Power OFF fu	Inction
7.4 Electrostatic mode		
8. Remote monitoring system	by a software	15 Page
8.1 Installation and connection of r	nonitoring program	
8.2 Monitoring program method		
8.3 Case of disconnecting to moni	toring program	
9. Attention		16 Page
10. Trouble shooting		17 Page
11. Warranty		18 Page

1. Feature



Welcome to become a customer of DIT.

Static electricity can cause lower production efficiency, such as damage on circuit pattern of products, particle deposition and etc.

"ARS-H002ZA" is the portable voltmeter, which can monitor and manage static electricity easily on production lines.

- 1) Detection of the electrostatic spot
- 2) Selecting the place for ionizer installation
- 3) Checking the Ion balance of ionizer and measuring static charge

Precision mode and Expansion mode

Precision mode detects static electricity from 0V to ± 2.00 kV, and Expansion mode detects it from ± 2.00 kV to ± 20.0 kV. Both modes are changed automatically during usage. Resolution is 10V from 0V to ± 9.9 9kV. When static electricity is under -10.0kV or +10.0kV, the resolution is 100V. In case of Ion Balance mode, it is able to be displayed in unit of 1V.

Rotary sensor head

ARS-H002ZA has a rotary head which makes convenient measurement in narrow space. Also a sensing plate for ion balance is designed with rotary integrated structure, so it is more convenient for use and storage of the product.

Real time Monitoring System by PC

It is able to check the measured data in real time and save the data via USB port. [Monitoring program is provided free of charge]



2. Specification

Dimension(Size)		134.0mm(L) x 68.6mm(W) x 26.0mm(H)		
Input Voltage		+9VDC ±10%(Alkaline battery use)		
Sensor		Oscillating Chopper Type (DIT)		
Measurement Dis	tance	25mm (±0.5mm) Fixed		
Measurement	Electrostatic mode	Precision mode : ±2kV (±10%) Expansion mode : ±20kV (±10%) (Auto change)		
Kange	Ion Balance mode	±300V (±10%)		
Decolution	Electrostatic mode	-9.99 ~ +9.99 kV : 10V -under 10.0kV or over +10.0kV : 100V		
Resolution	Ion Balance mode	1V		
Response Time		< 1sec.		
Function		Automatic power OFF, Alarm for power OFF, Zero setting, Electrostatic/Ion Balance mode, HOLD function		
Display		HTN-LCD[Digit and Gradation display]		
Material		Conductive ABS		
Operating time		Max approx. 30hours		
Communication		USB Communication		
Ambient Temperature		10℃ ~ 40℃ (50~104°F)		
Ambient Humidity		0~60%RH (Non-Condensing)		

* Specification can be changed without notice for performance improvement.

* This product is only designed for 25mm(±0.5mm) distance detection. It is available to detect for other distance, however performance is not guaranteed.

3. Package contents

Please check the package contents before use.

- ARS-H002ZA ------ 1ea Main body to detect static electricity/Ion Balance. LED of its head must be headed to the object.
- ② Battery(6LF22, 6LR61 9.0V Alkaline battery) ------ 1ea The product is operated after a battery is put in the back of main body.

(1) Battery socket is disconnected to prevent battery consumption during distribution. Please open the rear of the product and connect a battery to its socket.

- (2) In case of long-time storage, power could not be supplied by discharged battery. Please change a battery in this case.
- (3) Please check the specification when to change a battery.
- ③ USB CABLE ------ 1ea 1.2m USB cable is provided.
- ④ Earth CABLE ------ 1ea
 3m Earth cable is provided.
- Instruction Manual ------ 1ea
 Please read this instruction manual before using the product.



4. Parts name and functions

< Main body name and functions >

< Main body (front) >





NO	Name	Description
1	SENSING PLATE	Use for measuring ION BALANCE
2	LCD PANEL	Display measured value/Menu status/Power status/etc.
3	POWER BUTTON	Turn power ON/OFF
4	ZERO BUTTON	Set up to zero adjustment
5	HOLD BUTTON	Hold the measured value
6	IB BUTTON	Choose the Electrostatic mode or Ion Balance mode
7	LABEL 1	Caution
8	LABEL 2	Information of product and company
9	BATTERY BOX COVER	Cover of the battery slot

4. Parts name and functions



NO	Name	Description
1	U.A.R.T	External FIRMWARE terminal
2	USB Terminal	USB terminal for real time monitoring with PC
3	Ground Terminal	Earth Cable terminal
4	Plate rotary button	Both buttons should be pulled down when rotating the plate for measuring ion balance.
5	LED for focusing	Two LED guide lights help position the voltmeter at 25mm distance from a charged test object.
6	Sensing Hole	Sensor for static electricity

5. Dimension

< Front View >



< Top View >



< Side View >





6. Performance





ARS-H002ZA is a portable electrostatic voltmeter, which is affected by environment. Please read carefully the operating instruction below for an exact measurement.

7.1 Power supply[Battery use]

- 1) Battery socket is disconnected to prevent battery consumption during distribution. Please open the cover of battery slot and connect battery to its socket before use.
- 2) (Battery check)Power status is displayed on LCD as below. Please check the power status periodically and change a battery when the battery is out.



- 3) The procedure for changing battery is as below.
 - a cover of battery slot is on the back of main body. Pushing the cover, pull it down. Then it opens.
 - ② After taking a battery out of main body, disconnect the battery to its socket.
 - ③ Connect a new battery to the battery socket.
 - ④ Put the battery in main body. At this time, arrange wires of the socket inside the battery slot.
 - (5) Close the cover of the battery slot.

7.2 Measuring set-up

- 1) Connect main body to earthed objects with an ground cable.
- 2) It is recommended to wear Earth-Ring or ESD gloves to avoid a static electricity from human body.

Warning!! If Earth cable is not connected, inexact values could be measured. Please use the product after checking the product earthed.

- 3) Press [POWER] button to operate the product apart from a charged test objet or ionizer.
- 4) Check the display is same as below after power on.



[Initial DISPLAY on Electrostatic mode]



[Initial DISPLAY on Ion Balance mode]

Change a new battery



7.3 Setting zero point adjustment

In case that the initial value is not "0" though following the measuring set-up(7.2), Please set zero point adjustment as below.

- 1) Setting zero point adjustment on Electrostatic mode
 - (1) Position a sensing hole toward non-electrostatic area.
 - (2) After pressing [ZERO] button, check if the value is changed to "0" on display.
- Setting zero point adjustment on Ion Balance mode`
 - (1) Ion balance plate of head should be faced to the object.
 - (2) Touch an earthed wire on the plate, or remove remaining electric charge by finger touch.
 - (3) After pressing [ZERO] button, check if the value is changed to "0" on display.

- Warning! 1) If users set zero point adjustment during measurement, the measured value is not accurate by distortion of reference zero point.
 - 2) Setting zero point adjustment is not worked in below cases.([Zero] button is not available)
 - (1) Out of range : over ±0.80kV[Electrostatic mode] / over ±100V[Ion balance mode) (2) During HOLD function

7.4 Electrostatic mode

This product measures the voltage or electrostatic level on objects with an integral sensor on Electrostatic mode.

1) Check to set the electrostatic mode as below.



		Description
	1	Polarity of voltage
	2	Digit [-2kV~+2kV, -20kV~+20kV]
	3	Bar graph [approx. 2kV/1 bar] [L(-)/R(+)]
1	4	Unit of measure [kV]

2) Place the device at right angle to test object as below picture. Adjust the distance until two LED light mark become the one on the target.



LED	Measurement distance			
	Distance over 25mm			
	Right distance around 25mm			
9	Distance under 25mm			

Warning!!

An error of LED focus could occur according to measuring angle and material of the objects. Recommend to use aids such as ruler for exact measurement from 25mm distance.

3) Check the measured value and polarity of voltage on DISPLAY.



[ex1. measuring +1.25kV]



[ex2. measuring -12.5kV]

Warning!!

When "HHH" or "LLL" is displayed, the level of static electricity is over measurable range.
In this case please stop measuring because integral sensor can be damaged.
※ HHH : over (+)voltage measurable range[+20kV], LLL : over (-)voltage measurable range[-20kV]

7.5 Ion Balance mode

This product measures ion balance with sensing plate and an integral sensor in Ion Balance mode.

After pulling down the plate rotary button(①) on both sides of head, move the sensing plate(②) until above the sensing hole.





Warning!!

- 1) Make sure to pull down the plate rotary button on both sides of head when moving the sensing plate. Otherwise, fixing part for the sensing plate can be damaged.
 - 2) Before measuring, the sensing plate should be moved to above the sensing hole. Otherwise, the measured value could not be accurate.

	3
	ר
0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0 0	^⑤ [V]

	Description		
1	Ion balance mode		
2	Polarity of voltage		
3	Digit [-300V~+300V]		
4	Bar graph [approx. 30V/1 bar], [L(-)/R(+)]		
5	Unit of measure [V]		

3) How to position the sensing plate at ionizer to measure ion balance.

Please approach the sensing plate on ionizer to measure ion balance as much as distance between ionizer and object. Ion balance value could be different according to the distance.



- - 4) When measuring on Electrostatic mode, pull down the plate rotary button(1) on both side of head. Please fold the sensing plate(2) up for exposing the sensing hole.



7.6 HOLD function [Electrostatic mode/lon balance mode]

When the value on DISPLAY is hard to be checked, this function holds the value on DISPLAY. So it is convenient to check the value even after changing the position.

- 1) Follow the operating instruction for measurement.
- 2) Press [HOLD] button on main body once when to measure the value.
- 3) When positioning the product with measurable distance, the value is fixed and "HOLD" is displayed on LCD.



- 4) After checking the fixed value, press [HOLD] button once again. Then, "HOLD" is disappeared and the value and bar graph are changed in real time.
- Warning!!

 When the product is turned off on HOLD status, the fixed value is not saved.
 If measure other objects on HOLD status, the value is not displayed on LCD. Please measure after releasing HOLD status certainly.

7.7 Automatic Power OFF function [Electrostatic mode/Ion balance mode]

Power is turned off automatically with beep sound if not use for 5 minutes after measuring. This function can be set by [HOLD] button on main body.

- 1) This function is the default setting for extension of battery life.
- 2) If user want to turn off this function, press [HOLD] button over 3 seconds. Then, Automatic Power OFF function is turned off and it is displayed on LCD as below.



- 3) If not to use this function, the product keep continuously power-on until user turn off power. Please note the battery charge status.
- 4) If user want to reactivate this function, press [HOLD] button over 3 seconds. Then, blinking of the battery charge status stops and Automatic Power OFF function is reactivated.

Warning!!

- In case of deactivating Automatic Power OFF function, check the battery charge status because the product keep power on until user turns off by POWER button. Battery could be discharged.
- 2) When turning on the product with deactivating Automatic Power OFF function, this function is reactivated. If user want to deactivate it, repeat the steps above.

8. Remote monitoring system by a software

8.1 Installation and connection of monitoring program

- 1) Please download the USB driver and monitoring program on our web site.
- 2) Install the USB driver(FT232RL) to PC.
- 3) Connect ARS-H002ZA and PC with a provided USB cable.
- 4) Check the port number of driver. [available on Microsoft Management Console]

8.2 How to use monitoring program

ElectronicSen	sorPortableMonitor		3		4			5	- =
	Port COM1	Disconnect	MAX	21000	SET Displa	y Length	500	SET SAVEN	IONE
林東日接	Baud 9600	 EXIT 	MIN	-21000	SET Viev	v Style	Color SET	SAVEC	ж
	-8 Volt		MODE 2	!					
			Electr	onic Senso	r Monitor		Sensor Ok	2	
20000 -									
15000			_						
10000									
5000 -									
0 <u>=</u>									
-5000 -									
10000 <u>-</u> 									
15000									
20000 -									10.00
81	9:29:49.514 19:29:55.	272 19:30:00.934 1	9:30:06.595 19	:30:12.353 1 Tir	9:30:18.015 19:30: ne	23.773 19:30:	29.435 19:30:35	.097 19:30:40.855	19:30:40

- 1) Specify the same port as the port connecting PC and USB.
- 2) Set the communication speed. [Default setting : 9600]
- 3) Set range of the MAX, MIN voltage value on the Y-axis of the graph.
- 4) Set the display time on the X-axis of the graph.
- 5) Select whether to save measured value or not. When selecting save option, the graph output is not available.
 - * The measured value and time is saved continuously in excel file format.
- 6) After completing the settings, click "connect" button. Then the program starts.

8.3 Case of disconnecting to monitoring program

- 1) Check the status of USB cable connecting ARS-H002ZA and PC.
- 2) Check that the port number of USB driver is same as the port number of software.
- 3) Check to set 9600 baud on monitoring program settings.
- 4) Check to set 9600 baud on PC port settings.
- 5) Make sure if the measured value is over the range of the graph(Y: voltage/ X: time)
- 6) If problem is not solved, contact to the manufacturer or sales agents.

Warning!!

Monitoring program is provided for free. DIT dose not have any responsibility for damage by user, and customizing this program is not allowed.

9. Attention

ARS-H002ZA is a portable electrostatic voltmeter, which includes susceptible parts. Please read carefully the attention below before use.

- 1) After Reading the instruction manual, use the product according to instruction manual.
- 2) Make sure to keep the product away from water/oil/solvent/power/etc.
- 3) Do not use the product in place with corrosive gas such as acid/alkali/etc.
- 4) Do not give strong impact or vibration to the product. Sensor could be damaged.
- 5) Make sure to avoid a loaded heavy weight on the LCD panel.
- 6) Do not use the product in place with strong electromagnetic noise. It may cause malfunction of integral micro-chip.
- 7) Make sure to ground the product with a provided earthed cable before use.
- 8) The Sensing hole part should be avoided condensation.
- 9) Make sure that the sensing hole is not polluted by water and dust.
- 10) Do not touch or put sharp objects to the sensing hole.
- 11) When measuring ion balance, do not face the sensing hole to the object without the sensing plate.
- 12) Do not press buttons on main body with excessive force.
- 13) Keep the sensing plate at a humidity under 60%RH because of its high insulation. [recommended to use desiccator or plastic bag and desiccant]
- 14) The sensing plate should maintain its insulation. Otherwise, the measured value is not guaranteed. Please check the degree of insulation regularly as below.
 <How to check the degree of insulation for the sensing plate>
 - After 100V charging to the sensing plate, check the decreasing speed. It means that the insulation resistance is lower is low as much as the decreasing speed is faster.
 - (2) Measure the insulation resistance on the bridge of under the sensing plate. The insulation resistance should be over $10T\Omega$ ($10^{13}\Omega$).
- 15) Do not break or assembly the product. It may cause malfunction of the product. Also it is subject to restriction for exchanging or repairing the products.

10. Trouble shooting

Problems	Check points
Power is not supplied	 Check the battery connection Check the battery charge status If there is no problems above, contact to manufacturer or sales agents.
Unusual value on DISPLAY	 Check that ground cable is connected properly Check the settings for measurement mode
Error is bigger than real electrostatic value	 Check the measurement distance [25mm(±0.5mm)] Check the any other object affecting its performance. If there is no problems above, contact to manufacturer or sales agents.
LED light is OFF	1) Contact to manufacturer or sales agents
Burning smell during operation	1) Remove the battery and contact to manufacturer or sales agents
ETC.	1) Contact to manufacturer or sales agents

* If you cannot solve the problem with the above solution or if you have other problems not described above, please call manufacturer or leave a message on our website.

☞ DIT Quality Assurance Team TEL) +82-31-299-5464

Technology Market Solution

DIT Technology Application Experts are focused on delivering solutions for the Semiconductor, Electronic Assembly, Photovoltaic, Flat Panel Display, Disk Drive, Cleanroom. You can be sure of receiving maximum performance and reliability. We provide cost effective, innovative products best suited to your application needs.

Warranty

We, Dong II Technology Ltd. Manufactured this product under a strict quality control system and warrants it for1 year of period from the shipment date

However, we don't have any responsibility for

- 1) Any damage if the product is used in a way different from that is explained in this manual or remade by users arbitrarily.
- 2) Any damage led by improper usage. We recommend the installation circumstances in this manual, but that is just a recommendation and users are responsible for understanding the product's specification and judgment of suitability of usage.
- 3) Direct or indirect damage led by the product malfunction.